

Laboratory guidelines for radiometer calibration and characterisation

Ilmar Ansko, Viktor Vabson, Riho Vendt, Tartu Observatory, ESTONIA

Darmstadt 2022

Requirements for laboratories

Competent personnel

Environmental conditions

Measurement equipment

Calibration program

Traceability charts

Measurement procedures

Uncertainty budgets

Participation in comparison exercises

Everything listed above documented and monitored according to ISO 17025

TODO: self-declaration (wavelength, SNR, dark signal, linearity)

Requirements for measurements

Following the harmonized laboratory guidelines (D12)

Data in SI units with uncertainties according to ISO GUM

Traceability documents available

Sufficiently detailed measurement model


Cal/char reports shall meet ISO 17025 requirements

TODO: self-declaration (wavelength, SNR, dark signal, linearity)

List of cal/char tasks

1. Absolute calibration for radiometric responsivity
2. Long-term stability
3. Straylight and out of band response
4. Immersion factors (radiance, irradiance)
5. Angular response in air
6. Radiometric non-linearity
7. Accuracy of integration times
8. Dark signal
9. Thermal sensitivity
10. Polarization sensitivity
11. Temporal response
12. Wavelength scale
13. Signal-to-noise ratio
14. Pressure effects


The problem: small errors on top of large signals with drifts



Irradiance calibration



IOCCG: in close proximity
no differences

Reviewers: different lamp distances
baffling issues
back reflections
auxiliary certificates
diffuser's reference plane



Radiance calibration


IOCCG: same as previous
Reviewers: same as previous
accounting for panel BRDF



Stray light



IOCCG: Compatible
more application methods

Reviewers: double monochromator
pixel centroid
peak oversampling
entrance overfilling
suggesting tunable lasers
OOB importance



Immersion factors




IOCCG: unaltered
Reviewers: complaints about copy/paste
transmissivity of the collector



Angular response

IOCCG: compatible
added fixed azimuth requirement &
regular checking of reference signal

Reviewers: add D7 to references




Radiometric linearity

IOCCG

compatible
integration time method
possible thermal sensitivity
monochromator option

Reviewers

multiple lamp calibration
explanation of the monochromator method
adding uncertainty
distance vs. inttime method
dependent on the inttime?



Accuracy of integration times

| | |
|------------------|--|
| IOCCG | compatible no differences |
| Reviewers | none |

Dark signal

IOCCG

compatible


using the longest inttime

Reviewers

dark measurement policy

explanation of the exponential part


opaque pixels vs. shutter



Thermal sensitivity

IOCCG compatible
simultaneously with linearity,
dark signal, SNR, wavelength

Reviewers none



Polarization sensitivity

IOCCG compatible
regular check of the reference signal

Reviewers about application method




Temporal response

| | |
|-----------|------|
| IOCCG | TBD |
| Reviewers | none |

Wavelength scale

| | |
|------------------|--|
| IOCCG | compatible optional thermal sensitivity |
| Reviewers | pixel centroid yearly re-calibration table symbols |



Signal-to-noise ratio

IOCCG compatible
extended from signal properties
to the instrumental properties

Reviewers: none

Pressure effects

| | |
|------------------|-------------|
| IOCCG | TBD |
| Reviewers | none |

Gaps in the characterization guidelines

Out-of-band response

Accuracy of integration times

Temporal response

Pressure effects

General notes from reviewers

Re-calibration and re-characterization should depend on the uncertainty requirements

Some references missing

Add manufacturer's information to the re-calibration table

Explain/justify the ISO 17025 requirements

Expanding abbreviations

Missing illustrating graphs

Field exercise date and location

Naming of the companies/instruments

Paragraph & figure numbering

Comments in the CP files

Give full names of the related people

CP data formats

1. Absolute calibration (radiance, irradiance) + linearity + wavelength scale
2. Straylight
3. Angular response of irradiance sensors in air
4. Thermal sensitivity
5. Polarization sensitivity

Conclusions

Cal&char light sources differ from the natural ones; new source types needed

Instrument parameters (linearity, straylight, thermal, polarization) affect each other

Temporal & thermal drifts during long experiments shadow the small systematic effects

SNR with currently available sources limits the characterization wavelength range

The measurements are lengthy: software for automated lab measurements is needed

RAMSES: dark from opaque pixels not sufficient for some measurements